



बुटवल उपमहानगरपालिका

नगर कार्यपालिकाको कार्यालय

बुटवल, रुपन्देही
लुम्बिनी प्रदेश, नेपाल।



प.स.०८१/८२

च.न.

फोन : ०७१-५४०११२, ५४००३५

फ्याक्स : ०७१-५४१२८३

E-mail: info@butwalmun.gov.np

मिति: २०८२।०९।०७

दररेट उपलब्ध गराइदिने सम्बन्धी सूचना ।

प्रस्तुत विषयमा यस कार्यालयलाई फिजियोथेरापीसँग सम्बन्धित यसै साथ संलग्न स्पेशलिफिकेसन सहितका मालसामानहरुको प्रतिएकाई दररेट आवश्यक परेकोले सूचनाको मितिले सात दिन भित्र उक्त मालसामानहरुको भ्याट बाहेकको प्रतिएकाई दररेट उपलब्ध गराईदिनुहुन अनुरोध छ ।


मित्रमणि खनाल

प्रमुख प्रशासकीय अधिकृत

मित्रमणी खनाल

प्रमुख प्रशासकीय अधिकृत


BUTWAL SUB-METROPOLITAN CITY, BUTWAL RUPANDEHI


मित्रमणी खनाल
प्रमुख प्रशासकीय अधिकृत

Price Estimation for Physiotherapy Equipment,

S.N.	Name of Equipment/ Instruments/Tools	Description/Specifications	Unit Price	QTY	Total	Vat	Grand Total
1	Traction machine and kit with Bed -Lumbar Cervical (Smooth and Noiseless	Traction Force: 5 to 45 kg.Mode: Static and Intermittent. Hold Time: 20-80 sec. Rest Time: 1 Sec.-20 sec. Timer : 0-30 Minutes Mechanical Timer. Alarm: At the end of treatment and while activation of patient switch.Power Required: 220v & 50 Hz .Accessories: Cervical Collar, Cervical Spreader Bar, Lumbar Belt, Chest Belt and Ankle Traction Belt, ISO, CE or USFDA certified product.		1			
2	Interferential Therapy (IFT)-4 channel, vector mode and frequency (IFT Program:	Outputs: Two (4 Channels), Waveforms: Sine Wave, Symm Biphasic, DC, Monophasic, Modes : Interfential-Quadri polar with Vector, Premod; TENS- Continuous, Burst & Modulation; Galvanic- DC and Interrupt DC, Carrier Freq.- 4000 Hz (IFT) & Pulse Rate: 1 Hz - 250 Hz (Selectable), Pulse Width: 50 μ S - 400 μ S, Treatment Time: 5 - 60 Min (step in 5 min.), Output Voltage: 0 - 70 Vpp @ 500 Ω . ISO, CE and USFDA certified product.		1			
3	Therapeutic Ultrasound- Combined 3 MHZ and 1 MHZ (U. Sound frequency:	Frequency: 1 mhz and 3 mhz. Output mode: Continuous & pulsed. Treatment head: min. One(1). Treatment timer: 1 to 60 minutes. Selectable Duty Cycle and Frequency. % Duty Cycle-10%, 20% and 50%. Intensity: Continuous mode 0 to 2.5 watts/cm sq., pulsed mode: 0 to 3 watts/cm sq.		1			

		Accessories: Ultra-sound gel bottle. Power required: 220v. Ac, 50 hz), ISO, CE and USFDA certified product.					
4	TENS (Transcutaneous nerve stimulation)	4 body areas/ 2 patients can be treated at one time. Outputs: 4 Independent Outputs. Waveform: Sine Wave, Symm Biphasic, DC, Monophasic. Modes: Continuous, Fast Burst, Slow Burst, Ramp Burst, Wide Low Frequency, Narrow Low Freq. and Auto All. Frequency: 1 Hz - 250 Hz. Pulse Width: 50 μ S - 350 μ S (10 μ S / step). Timer: 1- 60 min., ISO, CE and USFDA certified product.		1			
5	Muscle stimulator	4 channel, pen and pad electrodes. Modes: IFT: Low, High, Wide & Fixed, TENS: Conv. Burst & Modulation, EMS: Conv. and Russian. IFT Carrier Freq. - 5000 Hz. Sweep: 1 - 10 Hz (Low) 80 - 150 Hz (High) and 1 - 250 Hz. (Wide) 10 - 200 Hz (Fixed). Pulse Width: TENS: 50 - 250 μ S and EMS: 200 - 450 μ S. Treatment Time: 1 - 60 Minutes. Power: Four 1.2 Volt AA Size Batteries and 5.5 Volt Adaptor., ISO, CE and USFDA certified product.		1			
6	Parallel bar	Almunium and stainless steel, ISO and CE certified product		1			
7	Parallel bar	wooden ramp for indoor use, ISO and CE certified product		1			
8	Quadriceps table	Standard quality, ISO and CE certified product		1			
9	Therabands	5 series set (all colors) Latex free quality, ISO and CE certified product		1			
10	Heel exercise	Metal design with waterproof wood, ISO and		1			


मित्रमणी खनाल
 प्रमुख प्रशासकीय अधिकारी

		CE certified product				
11	CPM machine knee	(Standard quality) operations :220V, 50HZ.Hold time:0 to 10 sec. dimensions: Approx.320mmx335mm x150 mm, ISO and CE certified product		1		
12	CPM machine Elbow	(standard quality) operations:220v AC, 50HZ, demensions: Approx.320mm x335mm x150mm, ISO and CE certified product		1		
13	Physio ball 55",65",90"	Standard quality, ISO and CE certified product		1		
14	Moist heat therapy unit (plus 4 standard heating pads)	Made from heavy gauge stainless steel sheet, double walled & well insulated in between. SPECIFICATIONS: Construction: S.S. Sheet double walled, well insulated. high. Heat Control: Thermostat to control Temperature from 30 to 110 deg. cen. with 2 Pilot Lamps. Heater: 1kw.specifically designed Heater. Hot Packs: 4, Accessories: Two forceps to get-out Hot-packs from hot water & Power Pin Plug. Finish: Polished finish inside, ISO and CE certified product		1		
15	Wax Bath Therapy unit (Wax including kit)	Construction: Double walled body, well insulated. Heaters: Two, 1kw each, special wax Heaters. Heat Control: Thermostat to control temperature from 00 to 80 deg. Cen. Indicator Lamps: Two, for Mains & Thermostat. Mounting: Mounted on Four casters. Power: 220-240 V AC. (Load, 2kw). Finish: Externally Powder Coated finish. Accessories: 6 Kg. Physio Wax, Power Plug, One brush for wax application, One Mug, One spatula & aluminum sheet mesh for		1		

		heaters, ISO and CE certified product					
16	Foot stepper	Standard quality, ISO and CE certified product		1			
17	Shoulder wheel	Standard quality, ISO and CE certified product		1			
18	Pulley set	Standard quality, ISO and CE certified product		1			
19	(Digital) Static Cycle exercise for cardio	Exercise bike with metal ABS, customized with capacity of 200-220 IBS, ISO and CE certified product		1			
20	Weight cuff set	Standard quality, ISO and CE certified product		1			
21	Dumbbell set	Standard quality, ISO and CE certified product		1			
22	Tilt table(electronic)	Lenth.200cm, height.50cm, width.65/70cm, maximum working load.150kg, under bed clearance.15cm, ISO and CE certified product		1			
23	Hand exercise table with springs, weights and bolts.	Hard wood table with laminate top. Fitted 6 pullys in a steel frame. Leather loops and nylon cord passing through the pulley with hanging weight underneath for fingers hand, wrist and forearm exercise, ISO and CE certified product		1			
24	Mobilization table/bed (low height for mat activity for neuro and Children)	Standard quality, ISO and CE certified product		1			

25	Laser Therapy Unit	Technical Feature: a. Lightweight and easy to use b. Frequency range: In Pulse Mode Frequency Range is 10 to 5000 Hz. c. Wavelength :850nm Central Wavelength of diode. d. Power: Single diode Applicator - 200mW. e. A range of Probe- single point & cluster f. Power input: 100-240Vac, 50/60Hz g. Treatment Time: 0-10 Minutes 2. Standard Accessories: a. Laser Main Unit b. laser probe- 1pc c. d. Set of Protective Goggles (Two pairs) e. Treatment Trolley f. Power Cable g. Manual 3. Power supply: 220 VAC \pm 10%, 50/60 Hz 4. Certificate of quality: ISO, CE or USFDA certified product		1			
26	Hand gripper exerciser	Local		1			
27	Exercise Treadmill- Hospital/Clinic use	Exercise Treadmill-Speed- 0.1-10 mph, 0-15 inclination, powered, various levels of programs, monitors heart rate, pulse, handrails), ISO and CE certified product		1			
28	Early Childhood Development Materials	Standard quality with different variety for sensory, touch, pressure etc.		1			
29	Floor matrrerss	Standard quality with 6 inch thickness, wrinkle free cover, water resistant.					
30	Poster correction mirror	wheeled metal design and wooden frame, wall mirror 4mm thick, dimensions : height 150 cm, width 80cm, with sandblasted posture correction net for corrective exercises.		1			
31	Assesment kit Sèt	Standard Quality with a set of goniometer & knee reflex hammer.		1			
32	Electrical heating pad	Standard quality		1			
33	Cold pad (gel filled)	Standard quality		1			

34	Children sensi exercise ball	Sensory Ball with spike surface(small-55)	1			
35	Children sensi exercise ball	Sensory Ball with spike surface(medium-65)				
36	Balance board -circle wooden	Standard Quality, Local	1			
37	Thermometer	Digital, Standard quality, ISO and CE certified product	1			
38	sphygmomanometer	Standard quality, ISO and CE certified product	1			
39	Anatomical chart-English	Standard quality	1			
40	stools-2ft steel made	Standard quality	1			
41	spirometer	Standard size, ISO and CE certified product	1			
42	pulse- oximeter	Standard quality, ISO and CE certified product	1			
43	Foam rollarfor mat activities	Standard quality	1			
44	Standing frame - child	Wooden/metal local made	1			
45	Standing frame - adult	Wooden/metal local made	1			
46	standing frame- child and adult	Metallic adjustable	1			
47	BP set	Best Quality, ISO and CE certified product	1			
48	Weight machine	Best Quality with Digital disply, ISO and CE certified product	1			
49	LED View Box (Single Film)	Single film LED view box, operates on mains electric supply, View box used for viewing the images of X-ray, CT/MRI at healthcare facilities, Ultra slim design, LED backlit and shall have separate on/off function with separate rotary continuous adjustable brightness control at the bottom of	1			

	panel for convenient operation, LED light source shall have at least 20000 hours of operation, Shall be able to hold one full large size CT/MR films at a time with film has maximum size of 43cmX 35cm, ISO and CE certified product.					
	TOTAL					